



SHETH VADILAL SARABHAI GENERAL HOSPITAL & SHETH CHINAI MATERNITY HOSPITAL,

Ellis Bridge, Ahmedabad-380 006.

MICROBIOLOGY DEPARTMENT

OOT SAMPLE

NO	SAMPLE	PRE FUMIGATION	POST FUMIGATION
1	Operation Table	Staph 'C' -Ve Isolated	No Bacteria Isolated
2	Operation Instrument Trolley	Gram Negative Bacilli Isolated	No Bacteria Isolated
3	Suction Machine	No Bacteria Isolated	No Bacteria Isolated
4	Suction Tube	No Bacteria Isolated	No Bacteria Isolated
5	Light	No Bacteria Isolated	No Bacteria Isolated
6	Floor	No Bacteria Isolated	No Bacteria Isolated
7	Wall	No Bacteria Isolated	No Bacteria Isolated

PLATE EXPOSED to AIR

Plate - 1 → NO BACTERIA ISOLATED
Plate - 2 → NO BACTERIA ISOLATED
Plate - 3 → NO BACTERIA ISOLATED
Plate - 4 → NO BACTERIA ISOLATED
Plate - 5 → NO BACTERIA ISOLATED

J. D. Ekdull
Professor & Head
Department of Microbiology,
Smt. M. H. L. Muni. Medical College,
Ellis Bridge, Ahmedabad-380 006

**BACTERIOLOGICAL
INCHARGE**



SHETH VADILAL SARABHAI GENERAL HOSPITAL & SHETH CHINAI MATERNITY HOSPITAL,

Ellis Bridge, Ahmedabad-380 006.

Date :-

MICROBIOLOGY DEPARTMENT

POST FUMIGATION STERILISATION EVALUTATION OF OOT

Unit of FOGSTAR was demonstrated in orthopedic operation theater on 27th august'2004. Swabs were taken before fumigation and after fumigation. 5 Plates were exposed to air. The culture report is enclosed. Sterilisation of OT was satisfactory.

Jyoti S. Deokule

Dr. Jyoti S. Deokule
Professor & Head
Department of Microbiology,
Smt. H. L. Juni. Medical College,
Ellis Bridge, Ahmedabad-380006